# Thermo Scientific Barnstead LabTower TII water purification system

The Barnstead LabTower TII converts tap water into high-purity water and provides water storage for a complete Type 2 water solution.



### **ROUTINE LABORATORY WORK**

- Rinsing lab glassware
- Supplying autoclaves and ultrapure water systems
- Preparing and diluting buffers, reagents, and media

#### **Consistent water quality**

- Type 2 laboratory water meets ASTM Type II, CSLS-CLRW and ISO 3696 international standards
- Designed to continuously recirculate purified reservoir water to preserve water quality even during periods of inactivity

#### **Compact mobile design**

- · Two system options with permeate flows of 20 or 40 L/hr
- Stand-alone design with integrated 100 L reservoir takes up zero bench space
- Castors allow easy relocation
- · Systems can be upgraded later to accommodate growing water demands

#### Two ways to dispense water

- Dispense directly from the system via the dispensing valve for calibration solutions, reagent preparation, filling of containers, general rinsing purposes, etc.
- Dispense from the reservoir to supply autoclaves, dishwashers, ultrapure water systems, etc.

#### Integrated 100 L reservoir

- Up to 100 L of high purity water ready for dispensing when needed
- · High purity polypropylene reservoir is opaque to light
- Conical bottom allows complete emptying for efficient cleaning and disinfection
- Adjustable setting for reservoir volume can be programmed for times when demand is low, allowing you to only store water that you can use

#### Clear information on system display

- The large four-line display is code-protected to prevent accidental changes in system settings
- Illuminated and easily readable, the display provides information on resistivity or conductivity, operating mode status, such as "production", "stand-by" or sanitization, and volume of the reservoir in %

#### **GLP-compliant documentation**

- Developed to meet or exceed GLP requirements
- Recorded and traceable data can be obtained by print-out via the RS-232 interface and accessory printer
- Highly qualified and precise measurement of the conductivity is ensured by the cell constant of 0.01 cm<sup>-1</sup> to an accuracy of  $\pm$  0.1°C

Pure Water Purification Systems

## pure water system with integrated 100 L tank

| Quick Look Comparison  |                        |                             |  |  |  |  |  |
|--|------------------------|-----------------------------|--|--|--|--|--|
|  | LabTower TII<br>system | LabTower TII + UV<br>system |  |  |  |  |  |
| Pure water production at 15°C, L/hr                              | 20 or 40               | 20 or 40                    |  |  |  |  |  |
| Resistivity at 25°C, MΩ·cm                                       | 10-1                   | 10-1                        |  |  |  |  |  |
| Conductivity, µS/cm  | 0.1-1                  | 0.1-1                       |  |  |  |  |  |
| Bacteria content, CFU/mL, with sterile filter                    | < 1                    | < 1                         |  |  |  |  |  |
| Particle content (0.2 $\mu\text{m}$ ) per mL with sterile filter | < 1                    | < 1                         |  |  |  |  |  |

#### Pretreatment

- 5 µm prefilter and activated carbon in the combi-filter cartridge protects the RO membrane from chlorine and particles
- Hardness stabilizing cartridge protects the reverse osmosis
  stage from hard water

#### **Reverse osmosis and reservoir**

- Pretreated water is pressure-forced through the permeable reverse osmosis module removing 97– 99% of all inorganic ions, 99% of dissolved organic substances as well as microorganisms and particles
- The built-in 100 L reservoir has a conical bottom outlet to optimize cleaning and sanitization

#### Ion exchange and UV

- R0 membrane removes up to 99% of impurities. All remaining ions are removed by the low TOC, high-purity resins in the ion exchange cartridge.
- Irradiation with UV light (TII units with UV option only) eliminates any bacteria or germs that are present

LabTower TII with optional hand dispenser

# Thermo Scientific Barnstead LabTower TII water purification system



### **Specifications and Ordering Information**

| Feed Water Specifications* |   |  |  |  |  |  |
|----------------------------|---|--|--|--|--|--|
| Source                     | Potable tap water softened or hardness stabilized   |  |  |  |  |  |
| Blocking index (SDI)       | < 3, with higher values, an upstream pretreatment<br>(model no. 09.4000) is to be installed |  |  |  |  |  |
| Conductivity, µS/cm        | < 1500  |  |  |  |  |  |
| Free chlorine, mg/L        | < 0.1   |  |  |  |  |  |
| Colloid index              | < 3   |  |  |  |  |  |
| ph-Range                   | 4-11  |  |  |  |  |  |
| Temperature, °C            | 2-35  |  |  |  |  |  |
| Pressure, psi (bar)        | 29-87 (2 -6 )   |  |  |  |  |  |

\*complete list of feed water specifications can be found in the operational manual.

|   |  |  | _           |                           | _   |                                       |                        |
|---|--|--|-------------|---------------------------|---|---------------------------------------|------------------------|
| LabTower TII  |  |  |             |                           |   |                                       |                        |
| Operating<br>pressure   | Electrical requirements  | Power consumption  |             | Feed water<br>connector V |   | Dimensions<br>W x D x H in (mm)       | Ambient<br>temperature |
| 29-87 psi<br>(2-6 bar)  | 90-240V,<br>50/60Hz  | 0.12kW   | 3           |                           |   | 17.7 x 22.8 x 59<br>450 x 580 x 1500) | 2-35°C                 |
| System Options  |  | UV option  |             | LabTo<br>TII :            |   | LabTower<br>TII 40                    | LabTower<br>TII 60     |
| <b>LabTower TII Systems*</b><br>All systems include an RO membrane, a high purity<br>water cartridge, 10 inch 5 μm prefilter with hardness<br>stabilizer cartridge, pressure regulator, 0.2 um filter,<br>integrated 100 L reservoir and UV lamp where applicable |  | System with UV   |             | 50132193                  |   | 50132141                              | 50132394               |
|   |  | System without UV  |             | 50132191                  |   | 50132196                              | 50132393               |
| Required Accessories  |  |  |             |                           |   |                                       |                        |
| LabTower TII Pretreatment<br>Both cartridges are required for complete pretreatment   |  | 5µm filter with activated ca<br>and a 10" filter housing | arbon       | 50134022                  |   | 50134022                              | 50134022               |
| Both our indges are required  |  | 1µm filter with a 10" filter                             | housing     | 09.4003                   |   | 09.4003                               | 09.4003                |
| Sterile vent filter for rese  | rvoir  |  |             | 50135142                  |   | 50135142                              | 50135142               |
| Sterile overflow for reser  | voir   |  |             | 501327                    | 714   | 50132714                              | 50132714               |
| Optional Accessories  |  |  |             |                           |   |                                       |                        |
| UV lamp assembly for reservoir, 230V, 50Hz only<br>Includes holder and UV lamp  |  | 06.5006  |             | 06.5006                   | 06.5006   |                                       |                        |
| Water watcher<br>Alerts the user to leaks. Available as 230V, 50Hz only   |  | 16.0129  |             | 16.0129                   | 16.0129   |                                       |                        |
| Printer<br>Utilizes RS-232 interface for safe documentation of all  |  | 120V, 50/60Hz  |             | AY1137X1                  |   | AY1137X1                              | AY1137X1               |
|   | measured values and faults with date and time in<br>compliance with GLP-Guidelines |  |             | 09.2207                   |   | 09.2207                               | 09.2207                |
| Qualification documents   | (IQ/OQ)  |  |             | I0QD0CE50134156           |   | I0QD0CE50134156                       | I0QD0CE50134156        |
| Hand Dispenser Kit<br>Hand dispenser with 3 meter cord that connects to tank. Ships with a 0.2 micron final filter.   |  | 50138221   |             | 50138221                  | 50138221  |                                       |                        |
| Replacement Consuma   | ables  |  |             |                           |   |                                       |                        |
| High purity water cartride  | High purity water cartridge  |  | 02.2850-LAB |                           | 02.2850-LAB   | 02.2850-LAB                           |                        |
| Reverse Osmosis membrane<br>LabTower TII 20 and 40 require two membranes, and LabTower TII 60 requires four<br>membranes, as indicated  |  | 22.00<br>(must ord                                       |             | 22.0087<br>(must order 2) | 22.0087 (must order 2)<br>and 22.0046 (must order 2 |                                       |                        |
| System UV-lamp  | System UV-lamp   |  | 09.4002     |                           | 09.4002   | 09.4002                               |                        |
| Sterile 0.2 µm filter   | Sterile 0.2 µm filter  |  | 09.1003     |                           | 09.1003   | 09.1003                               |                        |
| Cleaning solution   |  |  |             | 09.22                     | 02  | 09.2202                               | 09.2202                |
| oreaning solution   |  | North America/Latin Americ                               | ca          | CMX2                      | 25  | CMX25                                 | CMX25                  |
| Activated carbon cartridge with 5 µm prefilter  |  | 06.5201  |             | 06.5201                   | 06.5201   |                                       |                        |
| Hardness stabilizing cart   | Hardness stabilizing cartridge   |  |             | 06.5452                   |   | 06.5452                               | 06.5452                |
| Replacement UV lamp (re   | Replacement UV lamp (reservoir, 230V, 50Hz only)                                   |  |             | 09.50                     | 02  | 09.5002                               | 09.5002                |
| 5 µm prefilter and hardne   | 5 µm prefilter and hardness stabilizer   |  |             | 06.52                     | 04  | 06.5204                               | 06.5204                |
|   |  |  |             |                           |   |                                       |                        |

**TYPE 2** 

